

NEW BARRACKPORE MUNICIPALITY
New Barrackpore, North 24 Parganas

Memo No. NBM/PWD/ 1231/2024

Date :- 05/07/2024

EOI No. :- NBM/PWD/4/100 of 2024-2025

Date :- 05-07-2024

Chairperson /Executive Officer, New Barrackpore Municipality invites Expression of Interest (EOI) for selection of suitable rate for the following **Medical goods / products** from eligible & resourceful Companies / Firms/ Suppliers/ Agencies. Intending Bidders are requested to submit their application at their own letter head in writing / typing suitable rate against EOI Notice in a sealed envelope. It is not obligatory to accept any of the proposal about the matters of EOI Notice by the Agencies. After the end of the selection of EOI, the evaluation rate will be determined at the Quotation.

Sl. No.	Name of Parts	Quantity	unit	Suitable imposed rates with GST
1	<p>Name of work :-Required the good quality C-Arm Compatible Semi-Automatic Operating Table with Manual over ride mechanism:-</p> <p>Products Description :- i) Remote Operated Electric Table ii) The Base is Centric iii) Stainless fitting. iv) Radio - Translucent top. v) Base and cover are made of attractive, easy to clean stainless steel. vi) Sophisticated mechanics provide smooth, stepless articulation of the table top for precise patient positioning.</p> <p>Technical Details :- Height Adjustment (Max) ----700-750mm min and 1100-1150mm Max Side Tilt -----18°/20° Back Section ----- -40° to +80° Trendelenburg----- +25°/30° Reverse Trendelenburg----- 25°/40° Top sliding ----- 250mm to 30mm Max. width----- 530-580 mm with Rails Overall Length ----- 2000-2200 mm Flex/ Reflex ----- 220°/120° Kidney Bridge ----- 4 inches Power input to be 200-240V AC /50HZ fitted with Indian Plug</p>	1	No.	
2	<p>Name of work :- Required the Vertical Sterilizer 12" x 20" for Hospital</p> <p>Specification :- High pressure electrically heated vertical steam sterilizer for sterilizing of surgical instruments, dress material, linen, rubber, plastic material, injection liquids by means of steam under pressure of 15 to 21 psi. (Adjustable). Triple Chamber with inner chamber and steam jacket made of thick 18 SCW 304 AISI GRADE stainless steel. Outer Chamber made of 26 SGW Thick stainless- steel. The Lid is made 6MM of stainless- steel dye pressed and closed by wing nuts arrangement. Lid opens by Paddle Operating system. Joint less neoprene Gasket does not allow any leakage. fitted with safety valve & Relief valve, water level indicator, water inlet and drain valves. Supplied complete with cord and plug but without dressing bins. Suitable to work on 220 volts, 1 ph 50Hz, AC supply. 3 KW ISI Mark Heaters</p>	1	No.	

Last date of EOI submission of 12/07/2024 at 2.00 P.M.

For details visit Municipal Website www.nbmonline.org and Office Notice Board.

Executive Officer

New Barrackpore Municipality

Executive Officer
 New Barrackpore Municipality

Dr. A. Pal
 Medical Superintendent
 Dr. B.C. Roy General Hospital and
 Maternity Home